

# Calibrating Unit

## New Standard Solution Report

### Equipment

Alcotest 7110 MKIII-C  
Location: WALL TOWNSHIP POLICE  
Serial No.: ARXA-0037  
Calibration File No.: 01905  
Calib. Date: 04/17/2019  
Calib. No.: 00037  
Certification File No.: 01906  
Cert. Date: 04/17/2019  
Cert. No.: 00028  
Linearity File No.: 01907  
Lin. Date: 04/17/2019  
Lin. No.: 00028  
Solution File No.: 01936  
Soln. Date: 06/04/2019  
Soln. No.: 00360  
Sequential File No.: 01936  
File Date: 06/04/2019

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 19020  
Model No.: CU-34  
Serial No.: DDAB-0015  
Expires: 01/08/2021  
Bottle No.: 1018

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 10:42D        |                               |                           |
| Control 1 EC      | 0.100%         | 10:43D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.099%         | 10:43D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 10:44D        |                               |                           |
| Control 2 EC      | 0.097%         | 10:44D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.098%         | 10:44D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 10:45D        |                               |                           |
| Control 3 EC      | 0.097%         | 10:46D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.098%         | 10:46D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 10:46D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: PTL MCKAY #178

Badge No.: 178

Date: 06/04/2019