

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01601 Calib. Date: 04/17/2019 Calib. No.: 00031
Certification File No.: 01602 Cert. Date: 04/17/2019 Cert. No.: 00027
Linearity File No.: 01603 Lin. Date: 04/17/2019 Lin. No.: 00027
Solution File No.: 01644 Soln. Date: 06/15/2019 Soln. No.: 00307
Sequential File No.: 01644 File Date: 06/15/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0176

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:13D		
Control 1 EC	-.---%	20:14D	21.8°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	20:14D	21.8°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRADY

First Name: PATRICK

MI: R

Signature: Pat. Brady #175

Badge No.: 175

Date: 06/15/2019