

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01905
Certification File No.: 01906
Linearity File No.: 01907
Solution File No.: 01945
Sequential File No.: 01945

Serial No.: ARXA-0037

Calib. Date: 04/17/2019

Calib. No.: 00037

Cert. Date: 04/17/2019

Cert. No.: 00028

Lin. Date: 04/17/2019

Lin. No.: 00028

Soln. Date: 06/21/2019

Soln. No.: 00364

File Date: 06/21/2019

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 19020

Model No.: CU-34

Serial No.: DDAB-0015

Expires: 01/08/2021

Bottle No.: 1124

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:56D		
Control 1 EC	0.099%	08:56D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	08:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:57D		
Control 2 EC	0.097%	08:57D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:58D		
Control 3 EC	0.095%	08:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL. K #129

Badge No.: 129

Date: 06/21/2019