

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01601	Calib. Date:	04/17/2019
Certification File No.:	01602	Cert. Date:	04/17/2019
Linearity File No.:	01603	Lin. Date:	04/17/2019
Solution File No.:	01650	Soln. Date:	06/21/2019
Sequential File No.:	01650	File Date:	06/21/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	19020	Expires:	01/08/2021
		Bottle No.:	1287

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:56D		
Control 1 EC	0.099%	08:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:58D		
Control 2 EC	0.098%	08:58D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	08:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59D		
Control 3 EC	0.098%	09:00D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PK. K #129

Badge No.: 129

Date: 06/21/2019