

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01905 **Calib. Date:** 04/17/2019 **Calib. No.:** 00037
Certification File No.: 01906 **Cert. Date:** 04/17/2019 **Cert. No.:** 00028
Linearity File No.: 01907 **Lin. Date:** 04/17/2019 **Lin. No.:** 00028
Solution File No.: 01951 **Soln. Date:** 07/05/2019 **Soln. No.:** 00366
Sequential File No.: 01951 **File Date:** 07/05/2019

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0015
Control Solution %: 0.100% **Expires:** 04/09/2021
Solution Control Lot: 19110 **Bottle No.:** 0127

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:11D		
Control 1 EC	-.---%	08:12D	N/A	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	08:12D	N/A	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARINO

First Name: ANTHONY

MI: T

Signature: 

Badge No.: 181

Date: 07/05/2019