

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01601 **Calib. Date:** 04/17/2019 **Calib. No.:** 00031
Certification File No.: 01602 **Cert. Date:** 04/17/2019 **Cert. No.:** 00027
Linearity File No.: 01603 **Lin. Date:** 04/17/2019 **Lin. No.:** 00027
Solution File No.: 01669 **Soln. Date:** 07/08/2019 **Soln. No.:** 00311
Sequential File No.: 01669 **File Date:** 07/08/2019

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 04/09/2021
Solution Control Lot: 19110 **Bottle No.:** 0254

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:32D		
Control 1 EC	0.098%	01:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	01:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:33D		
Control 2 EC	0.097%	01:34D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:34D		
Control 3 EC	0.097%	01:35D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: _____

Badge No.: 165

Date: 07/08/2019