

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01905 Calib. Date: 04/17/2019 Calib. No.: 00037
Certification File No.: 01906 Cert. Date: 04/17/2019 Cert. No.: 00028
Linearity File No.: 01907 Lin. Date: 04/17/2019 Lin. No.: 00028
Solution File No.: 01958 Soln. Date: 07/18/2019 Soln. No.: 00371
Sequential File No.: 01958 File Date: 07/18/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 04/09/2021
Solution Control Lot: 19110 Bottle No.: 0258

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:05D		
Control 1 EC	0.098%	10:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06D		
Control 2 EC	0.096%	10:07D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08D		
Control 3 EC	0.095%	10:08D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: PTL. MCKAY #178

Badge No.: 178

Date: 07/18/2019