

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01601	Calib. Date:	04/17/2019
Certification File No.:	01602	Cert. Date:	04/17/2019
Linearity File No.:	01603	Lin. Date:	04/17/2019
Solution File No.:	01680	Soln. Date:	07/18/2019
Sequential File No.:	01680	File Date:	07/18/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	19110	Expires:	04/09/2021
		Bottle No.:	0377

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:55D		
Control 1 EC	0.099%	09:56D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:57D		
Control 2 EC	0.096%	09:57D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58D		
Control 3 EC	0.096%	09:59D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: PTL MCKAY #178

Badge No.: 178

Date: 07/18/2019