

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARXA-0037
Calibration File No.: 01905
Calib. Date: 04/17/2019
Calib. No.: 00037
Certification File No.: 01906
Cert. Date: 04/17/2019
Cert. No.: 00028
Linearity File No.: 01907
Lin. Date: 04/17/2019
Lin. No.: 00028
Solution File No.: 01972
Soln. Date: 08/27/2019
Soln. No.: 00375
Sequential File No.: 01972
File Date: 08/27/2019

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 19110
Model No.: CU-34
Serial No.: DDAB-0015
Expires: 04/09/2021
Bottle No.: 0136

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:44D		
Control 1 EC	0.097%	23:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	23:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:45D		
Control 2 EC	0.095%	23:45D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	23:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:46D		
Control 3 EC	0.095%	23:47D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: 

Badge No.: 174

Date: 08/27/2019