

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01601
Certification File No.: 01602
Linearity File No.: 01603
Solution File No.: 01722
Sequential File No.: 01722

Serial No.: ARZL-0146

Calib. Date: 04/17/2019
Cert. Date: 04/17/2019
Lin. Date: 04/17/2019
Soln. Date: 08/28/2019
File Date: 08/28/2019

Calib. No.: 00031
Cert. No.: 00027
Lin. No.: 00027
Soln. No.: 00318

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 19110

Model No.: CU-34

Serial No.: DDAB-0014
Expires: 04/09/2021
Bottle No.: 0372

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:05D		
Control 1 EC	0.096%	11:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:06D		
Control 2 EC	0.097%	11:07D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:08D		
Control 3 EC	0.095%	11:08D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: PTC MCKAY #178

Badge No.: 178

Date: 08/28/2019