

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01731	Calib. Date:	09/03/2019
Certification File No.:	01732	Cert. Date:	09/03/2019
Linearity File No.:	01733	Lin. Date:	09/03/2019
Solution File No.:	01751	Soln. Date:	09/29/2019
Sequential File No.:	01751	File Date:	09/29/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0891

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:11D		
Control 1 EC	0.097%	09:12D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	09:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:13D		
Control 2 EC	0.096%	09:13D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:14D		
Control 3 EC	0.096%	09:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARINO

First Name: ANTHONY

MI: T

Signature: 

Badge No.: 181

Date: 09/29/2019