

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01731	Calib. Date:	09/03/2019
Certification File No.:	01732	Cert. Date:	09/03/2019
Linearity File No.:	01733	Lin. Date:	09/03/2019
Solution File No.:	01756	Soln. Date:	10/08/2019
Sequential File No.:	01756	File Date:	10/08/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0890

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:06D		
Control 1 EC	0.097%	13:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	13:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:08D		
Control 2 EC	0.096%	13:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:09D		
Control 3 EC	0.096%	13:10D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PLK. K #129

Badge No.: 129

Date: 10/08/2019