

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01983 **Calib. Date:** 09/03/2019 **Calib. No.:** 00038
Certification File No.: 01984 **Cert. Date:** 09/03/2019 **Cert. No.:** 00029
Linearity File No.: 01985 **Lin. Date:** 09/03/2019 **Lin. No.:** 00029
Solution File No.: 02002 **Soln. Date:** 10/21/2019 **Soln. No.:** 00389
Sequential File No.: 02002 **File Date:** 10/21/2019

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-114
Control Solution %: 0.100% **Expires:** 06/10/2021
Solution Control Lot: 19150 **Bottle No.:** 0833

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:18D		
Control 1 EC	0.099%	03:19D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:20D		
Control 2 EC	0.096%	03:20D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:21D		
Control 3 EC	0.096%	03:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRADY

First Name: PATRICK

MI: R

Signature:

Pat. Brady #175

Badge No.: 175

Date: 10/21/2019