

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01731 **Calib. Date:** 09/03/2019 **Calib. No.:** 00032
Certification File No.: 01732 **Cert. Date:** 09/03/2019 **Cert. No.:** 00028
Linearity File No.: 01733 **Lin. Date:** 09/03/2019 **Lin. No.:** 00028
Solution File No.: 01762 **Soln. Date:** 10/21/2019 **Soln. No.:** 00325
Sequential File No.: 01762 **File Date:** 10/21/2019

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113
Control Solution %: 0.100% **Expires:** 06/10/2021
Solution Control Lot: 19150 **Bottle No.:** 1290

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:20D		
Control 1 EC	-.---%	01:20D	23.2°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	01:20D	23.2°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: _____



Badge No.: 177

Date: 10/21/2019