

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01983 Calib. Date: 09/03/2019 Calib. No.: 00038
Certification File No.: 01984 Cert. Date: 09/03/2019 Cert. No.: 00029
Linearity File No.: 01985 Lin. Date: 09/03/2019 Lin. No.: 00029
Solution File No.: 02008 Soln. Date: 10/25/2019 Soln. No.: 00392
Sequential File No.: 02008 File Date: 10/25/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 1287

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:38D		
Control 1 EC	0.098%	14:39D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	14:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:40D		
Control 2 EC	0.095%	14:40D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	14:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:41D		
Control 3 EC	0.095%	14:42D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

Badge No.: 129

Date: 10/25/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01983 Calib. Date: 09/03/2019 Calib. No.: 00038
Certification File No.: 01984 Cert. Date: 09/03/2019 Cert. No.: 00029
Linearity File No.: 01985 Lin. Date: 09/03/2019 Lin. No.: 00029
Solution File No.: 02007 Soln. Date: 10/25/2019 Soln. No.: 00391
Sequential File No.: 02007 File Date: 10/25/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 1287

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:45D		
Control 1 EC	0.098%	12:46D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:47D		
Control 2 EC	0.096%	12:47D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	12:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:48D		
Control 3 EC	0.094%	12:49D	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.098%	12:49D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	12:49D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL #129

Badge No.: 129

Date: 10/25/2019