

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
**Location:** WALL TOWNSHIP POLICE  
**Calibration File No.:** 01983 **Calib. Date:** 09/03/2019 **Calib. No.:** 00038  
**Certification File No.:** 01984 **Cert. Date:** 09/03/2019 **Cert. No.:** 00029  
**Linearity File No.:** 01985 **Lin. Date:** 09/03/2019 **Lin. No.:** 00029  
**Solution File No.:** 02012 **Soln. Date:** 10/30/2019 **Soln. No.:** 00393  
**Sequential File No.:** 02012 **File Date:** 10/30/2019

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDXB S3-114  
**Control Solution %:** 0.100% **Expires:** 06/10/2021  
**Solution Control Lot:** 19150 **Bottle No.:** 1286

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 07:38S        |                               |                           |
| Control 1 EC      | 0.098%         | 07:38S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.097%         | 07:38S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 07:39S        |                               |                           |
| Control 2 EC      | 0.095%         | 07:39S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.097%         | 07:39S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 07:40S        |                               |                           |
| Control 3 EC      | 0.096%         | 07:41S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.098%         | 07:41S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 07:42S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PKL #129

Badge No.: 129

Date: 10/30/2019