

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARXA-0037
Calibration File No.: 01983
Calib. Date: 09/03/2019
Calib. No.: 00038
Certification File No.: 01984
Cert. Date: 09/03/2019
Cert. No.: 00029
Linearity File No.: 01985
Lin. Date: 09/03/2019
Lin. No.: 00029
Solution File No.: 02015
Soln. Date: 11/02/2019
Soln. No.: 00394
Sequential File No.: 02015
File Date: 11/02/2019

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 19150
Model No.: CU-34
Serial No.: DDXB S3-114
Expires: 06/10/2021
Bottle No.: 1295

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:33S		
Control 1 EC	0.100%	05:33S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	05:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:34S		
Control 2 EC	0.096%	05:35S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	05:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:36S		
Control 3 EC	0.096%	05:36S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRADY

First Name: PATRICK

MI: R

Signature:

Pd. Brady #175

Badge No.: 175

Date: 11/02/2019