

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01983
Certification File No.: 01984
Linearity File No.: 01985
Solution File No.: 02016
Sequential File No.: 02016

Serial No.: ARXA-0037

Calib. Date: 09/03/2019
Cert. Date: 09/03/2019
Lin. Date: 09/03/2019
Soln. Date: 11/07/2019
File Date: 11/07/2019

Calib. No.: 00038
Cert. No.: 00029
Lin. No.: 00029
Soln. No.: 00395

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 19150

Model No.: CU-34

Serial No.: DDXB S3-114
Expires: 06/10/2021
Bottle No.: 1214

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:30S		
Control 1 EC	0.099%	11:31S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32S		
Control 2 EC	0.097%	11:32S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33S		
Control 3 EC	0.097%	11:34S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL. K. #129

Badge No.: 129

Date: 11/07/2019