

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C  
Location: WALL TOWNSHIP POLICE  
Serial No.: ARXA-0037  
Calibration File No.: 01983  
Calib. Date: 09/03/2019  
Calib. No.: 00038  
Certification File No.: 01984  
Cert. Date: 09/03/2019  
Cert. No.: 00029  
Linearity File No.: 01985  
Lin. Date: 09/03/2019  
Lin. No.: 00029  
Solution File No.: 02021  
Soln. Date: 11/10/2019  
Soln. No.: 00397  
Sequential File No.: 02021  
File Date: 11/10/2019

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 19150  
Model No.: CU-34  
Serial No.: DDXB S3-114  
Expires: 06/10/2021  
Bottle No.: 1230

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 20:30S        |                               |                           |
| Control 1 EC      | -.---%         | 20:30S        | N/A                           | SIMULATOR TEMP. ERROR     |
| Control 1 IR      | -.---%         | 20:30S        | N/A                           | SIMULATOR TEMP. ERROR     |
| Ambient Air Blank | -.---%         | --:--         |                               |                           |
| Control 2 EC      | -.---%         | --:--         | N/A                           |                           |
| Control 2 IR      | -.---%         | --:--         | N/A                           |                           |
| Ambient Air Blank | -.---%         | --:--         |                               |                           |
| Control 3 EC      | -.---%         | --:--         | N/A                           |                           |
| Control 3 IR      | -.---%         | --:--         | N/A                           |                           |
| Ambient Air Blank | -.---%         | --:--         |                               |                           |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MARINO

First Name: ANTHONY

MI: T

Signature: 

Badge No.: 181

Date: 11/10/2019