

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARXA-0037
Calibration File No.: 01983 Calib. Date: 09/03/2019 Calib. No.: 00038
Certification File No.: 01984 Cert. Date: 09/03/2019 Cert. No.: 00029
Linearity File No.: 01985 Lin. Date: 09/03/2019 Lin. No.: 00029
Solution File No.: 02024 Soln. Date: 11/14/2019 Soln. No.: 00399
Sequential File No.: 02024 File Date: 11/14/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 1294

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:58S		
Control 1 EC	0.098%	08:58S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	08:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59S		
Control 2 EC	0.096%	09:00S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00S		
Control 3 EC	0.095%	09:01S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

Badge No.: 129

Date: 11/14/2019