

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C
Location:	WALL TOWNSHIP POLICE

Serial No.: ARXA-0037

Location: WALL TOWNSHIP POLICE

Calib. No.: 00038

Cert. No.: 00029

Lin. No.: 00029

Soln. No.: 00400

Serial No.: DDXB S3-114

Expires: 06/10/2021

Bottle No.: 1294

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:58S		
Control 1 EC	0.098%	14:59S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00S		
Control 2 EC	0.094%	15:00S	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.098%	15:00S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	15:01S		
Control 3 EC	0.095%	15:02S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	15:02S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:03S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

MI: A

Badge No.: 129

Date: 11/14/2019