

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01769 **Calib. Date:** 12/13/2019 **Calib. No.:** 00034
Certification File No.: 01770 **Cert. Date:** 12/13/2019 **Cert. No.:** 00029
Linearity File No.: 01771 **Lin. Date:** 12/13/2019 **Lin. No.:** 00029
Solution File No.: 01775 **Soln. Date:** 12/23/2019 **Soln. No.:** 00329
Sequential File No.: 01775 **File Date:** 12/23/2019

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113
Control Solution %: 0.100% **Expires:** 06/10/2021
Solution Control Lot: 19150 **Bottle No.:** 0836

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:21S		
Control 1 EC	0.099%	15:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:22S		
Control 2 EC	0.099%	15:23S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:23S		
Control 3 EC	0.099%	15:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: PTL. MCKAY #478

Badge No.: 178

Date: 12/23/2019