

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01769 Calib. Date: 12/13/2019 Calib. No.: 00034
Certification File No.: 01770 Cert. Date: 12/13/2019 Cert. No.: 00029
Linearity File No.: 01771 Lin. Date: 12/13/2019 Lin. No.: 00029
Solution File No.: 01794 Soln. Date: 02/18/2020 Soln. No.: 00331
Sequential File No.: 01794 File Date: 02/18/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 0837

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:28S		
Control 1 EC	0.099%	11:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29S		
Control 2 EC	0.098%	11:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31S		
Control 3 EC	0.099%	11:32S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: RLK #129

Badge No.: 129

Date: 02/18/2020