

# Calibrating Unit

## New Standard Solution Report

|                         |                       |                         |
|-------------------------|-----------------------|-------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.: ARZL-0146   |
| Location:               | WALL TOWNSHIP POLICE  |                         |
| Calibration File No.:   | 01769                 | Calib. Date: 12/13/2019 |
| Certification File No.: | 01770                 | Calib. No.: 00034       |
| Linearity File No.:     | 01771                 | Cert. Date: 12/13/2019  |
| Solution File No.:      | 01810                 | Cert. No.: 00029        |
| Sequential File No.:    | 01810                 | Lin. Date: 12/13/2019   |
|                         |                       | Lin. No.: 00029         |
|                         |                       | Soln. Date: 05/10/2020  |
|                         |                       | Soln. No.: 00334        |
|                         |                       | File Date: 05/10/2020   |
| Calibrating Unit:       | WET                   | Model No.: CU-34        |
| Control Solution %:     | 0.100%                | Serial No.: DDXB S3-113 |
| Solution Control Lot:   | 19150                 | Expires: 06/10/2021     |
|                         |                       | Bottle No.: 1296        |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 09:43D        |                               |                           |
| Control 1 EC      | 0.099%         | 09:44D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.098%         | 09:44D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 09:44D        |                               |                           |
| Control 2 EC      | 0.098%         | 09:45D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.099%         | 09:45D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 09:46D        |                               |                           |
| Control 3 EC      | 0.099%         | 09:46D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.099%         | 09:46D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 09:47D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BAKER

First Name: JERRY

MI: A

Signature: \_\_\_\_\_

Badge No.: 185

Date: 05/10/2020