

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARXA-0037
Calibration File No.: 02030 Calib. Date: 02/24/2020 Calib. No.: 00040
Certification File No.: 02031 Cert. Date: 02/24/2020 Cert. No.: 00030
Linearity File No.: 02032 Lin. Date: 02/24/2020 Lin. No.: 00030
Solution File No.: 02041 Soln. Date: 06/12/2020 Soln. No.: 00406
Sequential File No.: 02041 File Date: 06/12/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 03/30/2022
Solution Control Lot: 20180 Bottle No.: 0943

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:24D		
Control 1 EC	0.101%	08:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:25D		
Control 2 EC	0.099%	08:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:26D		
Control 3 EC	0.100%	08:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: 

Badge No.: 129

Date: 06/12/2020