

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARXA-0037
Calibration File No.: 02030
Calib. Date: 02/24/2020
Calib. No.: 00040
Certification File No.: 02031
Cert. Date: 02/24/2020
Cert. No.: 00030
Linearity File No.: 02032
Lin. Date: 02/24/2020
Lin. No.: 00030
Solution File No.: 02042
Soln. Date: 07/02/2020
Soln. No.: 00407
Sequential File No.: 02042
File Date: 07/02/2020

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 20180
Model No.: CU-34
Serial No.: DDXB S3-114
Expires: 03/30/2022
Bottle No.: 1057

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:03D		
Control 1 EC	0.100%	07:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:04D		
Control 2 EC	0.098%	07:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:05D		
Control 3 EC	0.098%	07:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PKA129

Badge No.: 129

Date: 07/02/2020