

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02030 **Calib. Date:** 02/24/2020 **Calib. No.:** 00040
Certification File No.: 02031 **Cert. Date:** 02/24/2020 **Cert. No.:** 00030
Linearity File No.: 02032 **Lin. Date:** 02/24/2020 **Lin. No.:** 00030
Solution File No.: 02043 **Soln. Date:** 07/17/2020 **Soln. No.:** 00408
Sequential File No.: 02043 **File Date:** 07/17/2020

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-114
Control Solution %: 0.100% **Expires:** 03/30/2022
Solution Control Lot: 20180 **Bottle No.:** 1199

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:33D		
Control 1 EC	0.099%	11:33D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34D		
Control 2 EC	0.097%	11:35D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35D		
Control 3 EC	0.097%	11:36D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:37D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: 

Badge No.: 129

Date: 07/17/2020