

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01812 Calib. Date: 05/27/2020 Calib. No.: 00035
Certification File No.: 01813 Cert. Date: 05/27/2020 Cert. No.: 00030
Linearity File No.: 01814 Lin. Date: 05/27/2020 Lin. No.: 00030
Solution File No.: 01831 Soln. Date: 07/17/2020 Soln. No.: 00338
Sequential File No.: 01831 File Date: 07/17/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 0893

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:42D		
Control 1 EC	0.098%	11:43D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43D		
Control 2 EC	0.097%	11:44D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45D		
Control 3 EC	0.098%	11:45D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL K. 15 #129

Badge No.: 129

Date: 07/17/2020