

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02044 Calib. Date: 07/21/2020 Calib. No.: 00041
Certification File No.: 02045 Cert. Date: 07/21/2020 Cert. No.: 00031
Linearity File No.: 02046 Lin. Date: 07/21/2020 Lin. No.: 00031
Solution File No.: 02048 Soln. Date: 07/31/2020 Soln. No.: 00410
Sequential File No.: 02048 File Date: 07/31/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 03/30/2022
Solution Control Lot: 20180 Bottle No.: 1072

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:18D		
Control 1 EC	0.102%	09:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:19D		
Control 2 EC	0.100%	09:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	09:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:20D		
Control 3 EC	0.101%	09:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL KAZANOWSKY #129

Badge No.: 129

Date: 07/31/2020