

# Calibrating Unit New Standard Solution Report

|                         |                       |                         |
|-------------------------|-----------------------|-------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.: ARZL-0146   |
| Location:               | WALL TOWNSHIP POLICE  |                         |
| Calibration File No.:   | 01812                 | Calib. Date: 05/27/2020 |
| Certification File No.: | 01813                 | Cert. No.: 00035        |
| Linearity File No.:     | 01814                 | Cert. No.: 00030        |
| Solution File No.:      | 01847                 | Lin. No.: 00030         |
| Sequential File No.:    | 01847                 | Soln. No.: 00342        |
|                         |                       | File Date: 09/04/2020   |
| Calibrating Unit:       | WET                   | Model No.: CU-34        |
| Control Solution %:     | 0.100%                | Serial No.: DDAB-0015   |
| Solution Control Lot:   | 19150                 | Expires: 06/10/2021     |
|                         |                       | Bottle No.: 0838        |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 10:04D        |                               |                           |
| Control 1 EC      | 0.099%         | 10:05D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.098%         | 10:05D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 10:06D        |                               |                           |
| Control 2 EC      | 0.099%         | 10:06D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.098%         | 10:06D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 10:07D        |                               |                           |
| Control 3 EC      | 0.098%         | 10:08D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.098%         | 10:08D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 10:08D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

Badge No.: 129

Date: 09/04/2020