

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01812	Calib. Date:	05/27/2020
Certification File No.:	01813	Cert. Date:	05/27/2020
Linearity File No.:	01814	Lin. Date:	05/27/2020
Solution File No.:	01854	Soln. Date:	09/13/2020
Sequential File No.:	01854	File Date:	09/13/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0889

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:44D		
Control 1 EC	0.097%	03:45D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:45D		
Control 2 EC	0.097%	03:46D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	03:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:47D		
Control 3 EC	0.097%	03:47D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	03:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: 

Badge No.: 174

Date: 09/13/2020