

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01812	Calib. Date:	05/27/2020
Certification File No.:	01813	Cert. Date:	05/27/2020
Linearity File No.:	01814	Lin. Date:	05/27/2020
Solution File No.:	01861	Soln. Date:	10/12/2020
Sequential File No.:	01861	File Date:	10/12/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0993

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:21D		
Control 1 EC	0.098%	06:21D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:22D		
Control 2 EC	0.098%	06:23D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:23D		
Control 3 EC	0.099%	06:24D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	06:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: PTL. MCKAY #128

Badge No.: 178

Date: 10/12/2020