

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01812 Calib. Date: 05/27/2020 Calib. No.: 00035
Certification File No.: 01813 Cert. Date: 05/27/2020 Cert. No.: 00030
Linearity File No.: 01814 Lin. Date: 05/27/2020 Lin. No.: 00030
Solution File No.: 01863 Soln. Date: 10/16/2020 Soln. No.: 00347
Sequential File No.: 01863 File Date: 10/16/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 03/30/2022
Solution Control Lot: 20180 Bottle No.: 1073

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:19D		
Control 1 EC	0.099%	10:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:21D		
Control 2 EC	0.099%	10:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:22D		
Control 3 EC	0.099%	10:23D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

Badge No.: 129

Date: 10/16/2020