

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02064 **Calib. Date:** 10/27/2020 **Calib. No.:** 00042
Certification File No.: 02065 **Cert. Date:** 10/27/2020 **Cert. No.:** 00032
Linearity File No.: 02066 **Lin. Date:** 10/27/2020 **Lin. No.:** 00032
Solution File No.: 02070 **Soln. Date:** 11/25/2020 **Soln. No.:** 00418
Sequential File No.: 02070 **File Date:** 11/25/2020

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113
Control Solution %: 0.100% **Expires:** 06/10/2021
Solution Control Lot: 19150 **Bottle No.:** 0878

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:28S		
Control 1 EC	0.099%	11:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29S		
Control 2 EC	0.097%	11:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31S		
Control 3 EC	0.098%	11:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL K #129

Badge No.: 129

Date: 11/25/2020