

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01868 **Calib. Date:** 10/27/2020 **Calib. No.:** 00036
Certification File No.: 01869 **Cert. Date:** 10/27/2020 **Cert. No.:** 00031
Linearity File No.: 01870 **Lin. Date:** 10/27/2020 **Lin. No.:** 00031
Solution File No.: 01881 **Soln. Date:** 11/25/2020 **Soln. No.:** 00350
Sequential File No.: 01881 **File Date:** 11/25/2020

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-114
Control Solution %: 0.100% **Expires:** 03/30/2022
Solution Control Lot: 20180 **Bottle No.:** 0518

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:31S		
Control 1 EC	0.099%	11:31S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32S		
Control 2 EC	0.099%	11:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33S		
Control 3 EC	0.099%	11:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

Badge No.: 129

Date: 11/25/2020