

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02064
Certification File No.: 02065
Linearity File No.: 02066
Solution File No.: 02071
Sequential File No.: 02071

Serial No.: ARXA-0037

Calib. Date: 10/27/2020
Cert. Date: 10/27/2020
Lin. Date: 10/27/2020
Soln. Date: 12/10/2020
File Date: 12/10/2020

Calib. No.: 00042
Cert. No.: 00032
Lin. No.: 00032
Soln. No.: 00419

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 20180

Model No.: CU-34

Serial No.: DDXB S3-113
Expires: 03/30/2022
Bottle No.: 1066

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:18S		
Control 1 EC	0.099%	09:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:20S		
Control 2 EC	0.099%	09:20S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:21S		
Control 3 EC	0.099%	09:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

RLK #129

Badge No.: 129

Date: 12/10/2020