

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARZL-0146
Calibration File No.: 01868
Certification File No.: 01869
Linearity File No.: 01870
Solution File No.: 01900
Sequential File No.: 01900
Calib. Date: 10/27/2020
Cert. Date: 10/27/2020
Lin. Date: 10/27/2020
Soln. Date: 01/15/2021
File Date: 01/15/2021
Calib. No.: 00036
Cert. No.: 00031
Lin. No.: 00031
Soln. No.: 00354

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 20180
Model No.: CU-34
Serial No.: DDXB S3-114
Expires: 03/30/2022
Bottle No.: 0629

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:13S		
Control 1 EC	0.100%	11:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:15S		
Control 2 EC	0.099%	11:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:16S		
Control 3 EC	0.099%	11:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: Pl. K #129

Badge No.: 129

Date: 01/15/2021