

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01868	Calib. Date:	10/27/2020
Certification File No.:	01869	Cert. Date:	10/27/2020
Linearity File No.:	01870	Lin. Date:	10/27/2020
Solution File No.:	01923	Soln. Date:	03/20/2021
Sequential File No.:	01923	File Date:	03/20/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0432

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:45S		
Control 1 EC	0.098%	18:45S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:46S		
Control 2 EC	0.099%	18:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:47S		
Control 3 EC	0.099%	18:48S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: _____

Badge No.: 177

Date: 03/20/2021