

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01868 Calib. Date: 10/27/2020 Calib. No.: 00036
Certification File No.: 01869 Cert. Date: 10/27/2020 Cert. No.: 00031
Linearity File No.: 01870 Lin. Date: 10/27/2020 Lin. No.: 00031
Solution File No.: 01927 Soln. Date: 04/01/2021 Soln. No.: 00360
Sequential File No.: 01927 File Date: 04/01/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 0239

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:54S		
Control 1 EC	0.099%	08:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55S		
Control 2 EC	0.098%	08:56S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:57S		
Control 3 EC	0.099%	08:57S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:58S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

Badge No.: 129

Date: 04/01/2021