

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02072 **Calib. Date:** 04/15/2021 **Calib. No.:** 00043
Certification File No.: 02073 **Cert. Date:** 04/15/2021 **Cert. No.:** 00033
Linearity File No.: 02074 **Lin. Date:** 04/15/2021 **Lin. No.:** 00033
Solution File No.: 02076 **Soln. Date:** 05/16/2021 **Soln. No.:** 00421
Sequential File No.: 02076 **File Date:** 05/16/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 03/30/2022
Solution Control Lot: 20180 **Bottle No.:** 1058

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:45D		
Control 1 EC	0.100%	06:45D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:45D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:46D		
Control 2 EC	0.098%	06:47D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:47D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:47D		
Control 3 EC	0.098%	06:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	06:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MAZZACCARO

First Name: VINCENT

MI: J

Signature: Sgt. Vincent Mazzaccaro #172

Badge No.: 172

Date: 05/16/2021