

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01932	Calib. Date: 04/15/2021
Certification File No.:	01933	Cert. No.: 00037
Linearity File No.:	01934	Cert. Date: 04/15/2021
Solution File No.:	01957	Lin. No.: 00032
Sequential File No.:	01957	Lin. Date: 04/15/2021
		Soln. No.: 00364
		File Date: 05/28/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	20180	Expires: 03/30/2022
		Bottle No.: 0633

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:51D		
Control 1 EC	0.099%	10:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:52D		
Control 2 EC	0.099%	10:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:53D		
Control 3 EC	0.099%	10:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL K #129

Badge No.: 129

Date: 05/28/2021