Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE				
Calibration File No.:	01932	Calib. Date:	04/15/2021	Calib. No.:	00037
Certification File No.:	01933	Cert. Date:	04/15/2021	Cert. No.:	00032
Linearity File No.:	01934	Lin. Date:	04/15/2021	Lin. No.:	00032
Solution File No.:	01970	Soln. Date:	06/25/2021	Soln. No.:	00366
Sequential File No.:	01970	File Date:	06/25/2021		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAB-0015
Control Solution %:	0.100%			Expires:	03/30/2022
Solution Control Lot:	20180			Bottle No.:	0637

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:21D		
Control 1 EC	0.099%	11:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:22D		
Control 2 EC	0.098%	11:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23D		
Control 3 EC	0.099%	11:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY First Name: PTL MI: A

Signature: P.L. K #129

Badge No.: 129

Date: 06/25/2021