

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01932 Calib. Date: 04/15/2021 Calib. No.: 00037
Certification File No.: 01933 Cert. Date: 04/15/2021 Cert. No.: 00032
Linearity File No.: 01934 Lin. Date: 04/15/2021 Lin. No.: 00032
Solution File No.: 02001 Soln. Date: 08/01/2021 Soln. No.: 00369
Sequential File No.: 02001 File Date: 08/01/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 03/30/2022
Solution Control Lot: 20180 Bottle No.: 0407

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:46D		
Control 1 EC	0.098%	16:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:48D		
Control 2 EC	0.098%	16:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:49D		
Control 3 EC	0.097%	16:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HUDSON

First Name: JUSTIN

MI:

Signature: _____

Justin Hudson
A/G

Badge No.: 161

Date: 08/01/2021