

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02072 **Calib. Date:** 04/15/2021 **Calib. No.:** 00043
Certification File No.: 02073 **Cert. Date:** 04/15/2021 **Cert. No.:** 00033
Linearity File No.: 02074 **Lin. Date:** 04/15/2021 **Lin. No.:** 00033
Solution File No.: 02084 **Soln. Date:** 09/05/2021 **Soln. No.:** 00425
Sequential File No.: 02084 **File Date:** 09/05/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 03/30/2022
Solution Control Lot: 20180 **Bottle No.:** 1060

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:03D		
Control 1 EC	0.098%	10:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04D		
Control 2 EC	0.096%	10:05D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:05D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06D		
Control 3 EC	0.097%	10:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BAKER First Name: JERRY

Signature: P.H. J.B. #185

Badge No.: 185

Date: 09/05/2021

MI: A