

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01932	Calib. Date:	04/15/2021
Certification File No.:	01933	Cert. No.:	00037
Linearity File No.:	01934	Cert. Date:	04/15/2021
Solution File No.:	02014	Lin. No.:	00032
Sequential File No.:	02014	Soln. Date:	09/05/2021
		Soln. No.:	00370
		File Date:	09/05/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	20180	Expires:	03/30/2022
		Bottle No.:	0291

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:47D		
Control 1 EC	0.099%	09:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:48D		
Control 2 EC	0.098%	09:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:49D		
Control 3 EC	0.098%	09:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:50D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUHL

First Name: FRANK

MI: A

Signature:  159

Badge No.: 159

Date: 09/05/2021