

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02072
Certification File No.: 02073
Linearity File No.: 02074
Solution File No.: 02085
Sequential File No.: 02085

Serial No.: ARXA-0037

Calib. Date: 04/15/2021
Cert. Date: 04/15/2021
Lin. Date: 04/15/2021
Soln. Date: 10/10/2021
File Date: 10/10/2021

Calib. No.: 00043
Cert. No.: 00033
Lin. No.: 00033
Soln. No.: 00426

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 20460

Model No.: CU-34

Serial No.: DDAB-0014
Expires: 10/21/2022
Bottle No.: 1388

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:38D		
Control 1 EC	0.099%	00:39D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.097%	00:39D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:39D		
Control 2 EC	0.097%	00:40D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.097%	00:40D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:41D		
Control 3 EC	0.097%	00:41D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:41D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARINO

First Name: ANTHONY

MI: T

Signature: 

Badge No.: 181

Date: 10/10/2021