

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01932 **Calib. Date:** 04/15/2021 **Calib. No.:** 00037
Certification File No.: 01933 **Cert. Date:** 04/15/2021 **Cert. No.:** 00032
Linearity File No.: 01934 **Lin. Date:** 04/15/2021 **Lin. No.:** 00032
Solution File No.: 02019 **Soln. Date:** 10/10/2021 **Soln. No.:** 00371
Sequential File No.: 02019 **File Date:** 10/10/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0015
Control Solution %: 0.100% **Expires:** 10/21/2022
Solution Control Lot: 20460 **Bottle No.:** 1393

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:33D		
Control 1 EC	0.100%	00:34D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:35D		
Control 2 EC	0.099%	00:35D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:36D		
Control 3 EC	0.099%	00:37D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARINO

First Name: ANTHONY

MI: T

Signature: 

Badge No.: 181

Date: 10/10/2021