

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02021 **Calib. Date:** 10/13/2021 **Calib. No.:** 00038
Certification File No.: 02022 **Cert. Date:** 10/13/2021 **Cert. No.:** 00033
Linearity File No.: 02023 **Lin. Date:** 10/13/2021 **Lin. No.:** 00033
Solution File No.: 02030 **Soln. Date:** 11/07/2021 **Soln. No.:** 00373
Sequential File No.: 02030 **File Date:** 11/07/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-114
Control Solution %: 0.100% **Expires:** 10/21/2022
Solution Control Lot: 20460 **Bottle No.:** 1395

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:41S		
Control 1 EC	0.098%	04:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:43S		
Control 2 EC	0.097%	04:43S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:44S		
Control 3 EC	0.098%	04:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:46S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CACCIATORE

First Name: MICHAEL

MI: J

Signature: 

Badge No.: 187

Date: 11/07/2021