

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARXA-0037
Calibration File No.: 02086
Calib. Date: 10/13/2021
Calib. No.: 00044
Certification File No.: 02087
Cert. Date: 10/13/2021
Cert. No.: 00034
Linearity File No.: 02088
Lin. Date: 10/13/2021
Lin. No.: 00034
Solution File No.: 02092
Soln. Date: 11/12/2021
Soln. No.: 00429
Sequential File No.: 02092
File Date: 11/12/2021
Calibrating Unit: WET
Model No.: CU-34
Serial No.: DDXB S3-113
Control Solution %: 0.100%
Expires: 10/21/2022
Solution Control Lot: 20460
Bottle No.: 1466

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:37S		
Control 1 EC	0.099%	12:38S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:38S		
Control 2 EC	0.097%	12:39S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:40S		
Control 3 EC	0.096%	12:40S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: PTL. MCKAY #178

Badge No.: 178

Date: 11/12/2021