

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARZL-0146
Calibration File No.: 02021 Calib. Date: 10/13/2021 Calib. No.: 00038
Certification File No.: 02022 Cert. Date: 10/13/2021 Cert. No.: 00033
Linearity File No.: 02023 Lin. Date: 10/13/2021 Lin. No.: 00033
Solution File No.: 02039 Soln. Date: 12/03/2021 Soln. No.: 00376
Sequential File No.: 02039 File Date: 12/03/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 03/30/2022
Solution Control Lot: 20180 Bottle No.: 0408

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:24S		
Control 1 EC	0.101%	16:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:26S		
Control 2 EC	0.100%	16:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:27S		
Control 3 EC	0.101%	16:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: AL. MCKAY #178

Badge No.: 178

Date: 12/03/2021